

# EPSILON

passion to lead

EPSILON  Solution



Certified & approved By:



# Technical Specifications

SPECIFICATIONS	HIGH FREQUENCY C-ARM SYSTEMS	
	EP VICTORY 3.5 Advance	EP VICTORY RA6
MODEL	EP VICTORY 3.5 Advance	EP VICTORY RA6
Max Output (KW)	3.5	6
Fluroscopy KVMax.	110	120
Fluroscopy mA Max.	3.5	5
Radiography KV Max.	110	120
Radiography mA Max.	80	100
Tube Head	Stationary	Rotating
Focal Spot	Dual	
Focal Spot Size (mm)	0.6 & 1.5	0.3 & 6
Image System	Digital	
I. I. Tube	9" (230 mm) Triple Field (Normal - 9", Zoom <sub>1</sub> - 6", Zoom <sub>2</sub> - 4.5")	
LCD / LED Monitor	2 Nos. (17" / 19")	
Memory	100 Frames / 32 Frames / 100 Frames USB / LAN Support / Digital Image Processor	
ABS Facility	Available	
Fluroscopy	Continuous	Continuous + Pulse
Stabilizer	Optional	
Cassette Holder	Standred	
Power Required	1Ø 190-240 VAC 50 Hz with 15 Amps Power Socket (Required Proper Line Earthing)	

The company reserve the rights to add / delete / modify in details of specifications and appearance mentioned here in.



## Epsilon Healthcare Solutions (P) Ltd.

Office : 103-104, Amar Vijay Complex, Hotel Man Singh Lane,  
S. C. Road, Jaipur-01, Raj. (INDIA), Tel : +91.141.2362854

Works : G-1-57, RIICO Industrial Area Extension, Phase-2, Bagru, Jaipur, Raj. (INDIA)

E-mail : marketing@epsilonhealthcare.com, Web : www.epsilonhealthcare.com

Marketed by:

**Toll Free : 1800 1234 122**