

EPSILON

passion to lead

Certified & approved By:



EP-100

COUNTER BALANCE

Diagnostic Medical X-ray System

Technical Specifications

PARAMETER	Description
Type of Generator	2 pulse, Full wave rectified
Maximum Output	10 KVA / 8 KW
mA Range	25-100mA
KV Selection	30 to 100 kV
Exposure Time	0.05 to 12 Sec.
Input Supply	230 V AC, 1 Phase
Digital Display	mA, mAs & kV Selection
Control	APR / Manual
X-Ray Tube	DSA-3 OR XD-3
Tube Type	Stationary Anode
Equipment Approvals	BIS, ISO : 9001, 13485, AERB, CE

Specifications and appearance are subject to change without prior notice.

OPTIONS

APR Control



Manual Control



Binder • Foot Rest • Vertical Bucky • Motorized Table • Cones & Accessories

Epsilon Healthcare Solutions (P) Ltd.

Office : 103-104, Amar Vijay Complex, Hotel Man Singh Lane, S. C. Road, Jaipur-01, Raj. (INDIA), Tel : +91.141.2362854

Works : G-1-57, RIICO Industrial Area Extension, Phase-2, Bagru, Jaipur, Raj. (INDIA), Tel : +91.141.2168318

E-mail : marketing@epsilonhealthcare.com, Web : www.epsilonhealthcare.com

Distributor Stamp: